



## AUTHORIZATION TO MARK

This authorizes the application of the Certification Marks shown below to the Product Covered (also to the multiple listee model identified on the correlation page of the Listing Report where applicable) when made in accordance with the Description and under the conditions set forth in the Certification Agreement and Listing Report:

**Applicant:** Anton Paar GmbH  
Anton-Paar Str. 20  
8054 Austria

**Contact:** Name: Mr. Pöss

Phone: +43 316 257 - 475

**Manufacturer:** *same as applicant*

**Party Authorized To Apply Mark:** *Same as Manufacturer*

**Report Issuing Office:** *Intertek Deutschland GmbH, Innovapark 20, 87600 Kaufbeuren*

**Report No.:** 03KFI001575-ETL

**Product Covered:** Microwave Reaction System, Model Multiwave 3000

**Description:** *Microwave for sample preparation*

**Standard(s):** Standard for Electrical Equipment for Laboratory Use UL61010A-1, 1<sup>st</sup> Ed., dated January 2002, Part 2: Particular Requirements for Laboratory Equipment for the Heating of Materials UL61010-2-010, 1<sup>st</sup> Edition, dated March 13, 2002 and the Standard for Safety Requirements for electrical equipment for measurement, control and laboratory use CAN/CSA C22.2 No. 1010-1-92, 1<sup>st</sup> Ed. dated Dec.1992 + A2:1995 and CAN/CSA C22.2 No. 1010-2-10, 1<sup>st</sup> Edition and IEC 61010-2-081:2001+A1:2003 Particular Requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes.

This document is the property of Intertek Testing Services and is not transferable. Only the Applicant may reproduce this document. The certification mark(s) may be applied only at the above noted location of the Party Authorized To Apply Mark.



Authorized by: \_\_\_\_\_

*Jonas Backstad*  
Jonas Backstad for  
William T. Starr, Certification Manager

Date: 19 September 2005

Control Number: 3047470

Intertek Testing Services NA Inc.  
165 Main Street, Cortland, NY 13045  
Telephone 800-345-3851 or 607-753-6711 Fax 607-756-6699